



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Quality

Product Name: PKS-ATP Positive Control Tablets

Lot Number: PKS-ATP-022812B

Expiration Date: September 2016

Order Codes: ATP-PLUS-4, ATP- PLUS -20

This product is made from Sigma Adenosine 5'-triphosphate disodium salt, Product Number A2383, and contains $6.0 \pm 1.5 \times 10^3$ femto moles per tablet. Finished product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

PocketSwab Plus Positive Control results of LUM channel for novaLUM should be 10,000 to 40,000 RLU. Results for Charm FireFly and Charm II 6600 or 7600 should be 1,000 to 10,000 RLU. Results for the FireFly2 should be from 3,000 to 20,000 RLU.

WaterGiene Positive Control results of H2OG channel for novaLUM should be greater than 50,000,000. Results for LUM channel of novaLUM should be 130,000 to 200,000 RLU. Results for Charm FireFly should be 35,000 to 55,000 RLU. Results for the FireFly2 should be from 70,000 to 100,000 RLU.

A handwritten signature in black ink, appearing to read 'Sam Atoui', is written over a horizontal line.

Name: Sam Atoui
Title: QA Documentation Manager
Date: March 17, 2015

Nothing works like a Charm