CERTIFICATE OF ANALYSIS

PRODUCT:

100 ml Zero Control Standard

LOT NUMBER:

ZCS-015

EXPIRATION DATE:

June 2018

ORDER CODE:

ZCS, ZCS-10, ZCS-20

DESCRIPTION:

Zero Control Standard is certified organic whole milk powder from bovines of USA only. When reconstituted as indicated on the label (with 100 ml of 40°C water), it is comparable, within 10% of antimicrobial drug-free and

aflatoxin-free cow's milk.

Zero Control Standard is qualified as antimicrobial drugfree and aflatoxin-free by the Charm II Tests for Beta-

lactams, Sulfa Drugs, Tetracyclines, Macrolides,

Aminoglycosides, Novobiocin, Chloramphenicol and Other Amphenicols, and Aflatoxins, and ROSA MRL Beta-lactam

Test.

Finished product quality control approval is based on the Charm II Tests, the ROSA MRL Beta-lactam Test, the Charm *B.stearothermophilus* Tablet Disc Assay, and the broad spectrum microbial inhibition Charm CowSide II and Blue-Yellow II Tests. Shelf life studies, heat stress studies

and frozen stability studies are performed to verify

operator's manual claims.

PREPARATION:

See operator's manual.

STORAGE:

See operator's manual.

PREPARED BY:

David R. Legg

SIGNATURE:

TITLE:

Vice President, Quality Assurance

Paul R Legg

DATE:

October 05, 2016