

***B. stearo*thermophilus Spore Tablet  
Colony Count Certification**

Lot number: BSTTBL-030

Expiration date: April 30, 2024

Certified colony count: 94 x 10<sup>6</sup> CFU/tablet

Certification date: May 3, 2023

Colony count data supporting certification (100 x 10<sup>6</sup> CFU/tablet).

Tablet #	Colony Counts			Mean
1	123	85	106	105
2	125	97	109	110
3	99	102	101	101
4	101	106	85	97
5	84	94	69	82
6	84	87	74	82
7	83	86	108	92
8	86	79	90	85
9	112	92	96	100
10	104	81	75	87
			<b>Mean</b>	<b>94</b>
				<b>x 10<sup>6</sup> CFU/tablet</b>

Colony counts of each lot are certified to be 90 x 10<sup>6</sup> - 110 x 10<sup>6</sup> CFU/tablet.

The average colony count of 10 tablets will be within 10% of the certified colony count (above) for that lot.



Name: David R. Legg

Title: Vice President, Quality Assurance

Date: May 3, 2023



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Quality

Product Name: B. Stearothermophilus Spore Tablets

Lot Number: BSTTBL030

Expiration Date: April 2024

Order Code: Spore

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

The Spore tablets contain *Geobacillus stearothermophilus* var. *calidolactis*.

A handwritten signature in black ink, appearing to read "Sam Atoui", is written above a horizontal line.

Name: Sam Atoui

Title: Quality Assurance Documentation Manager

Date: 05/04/2023