



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Quality

Product Name: Charm MRLBLTET2A Test for Milk

Lot Number: MRLBLTET2A-006O

Expiration Date: October 2023

Order Codes: LF-MRLBLTET2A

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications

A handwritten signature in black ink, appearing to read "Sam Atoui", is written above a horizontal line.

Name: Sam Atoui

Title: Quality Assurance Documentation Manager

Date: 03/16/2023

TYPICAL COUNTS for
Charm MRL Beta-lactam and Tetracycline 2 Minute Test

See Operator's Manual for use and application.

Kit Code: LF-MRLBLTET2A

Lot Number: MRLBLTET2A-006O

Charm ROSA Reader

Negative Control	-2395	2716	MRLBLTET/TET Positive Tablet 4 ppb Penicillin G and 100 ppb Oxytetracycline
	-1940	2898	
	-2341	2411	
	-2836	2917	
	-2559	3304	
	-2779	2966	
	-2212	2611	
	-2593	2839	
	-2320	2869	
	-2460	2482	

Note: Typical counts are provided as a performance guide only.
Operators' daily control numbers may be compared to the supplied
information as an operator, equipment and analyzer calibration check. Daily
control data should typically test in the range of data points provided.
The control point for the MRL Beta-lactam and Tetracycline 2 Minute Test is always zero.