



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Quality

Product Name: Charm MRLBLTET2A Test for Milk

Lot Number: MRLBLTET2A-007A

Expiration Date: January 2024

Order Codes: LF-MRLBLTET2A

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications

A handwritten signature in black ink, appearing to read "Sam Atoui", is written above a horizontal line.

Name: Sam Atoui

Title: Quality Assurance Documentation Manager

Date: 06/08/2023

TYPICAL COUNTS for
Charm MRL Beta-lactam and Tetracycline 2 Minute Test

See Operator's Manual for use and application.

Kit Code: LF-MRLBLTET2A

Lot Number: MRLBLTET2A-007A

Charm ROSA Reader

Negative Control	-2226	2739	MRLBLTET/TET Positive Tablet 4 ppb Penicillin G and 100 ppb Oxytetracycline
	-2090	1748	
	-2138	2470	
	-2234	2937	
	-2503	2349	
	-1868	2505	
	-2174	3070	
	-2291	2156	
	-1787	2152	
	-1974	2139	

Note: Typical counts are provided as a performance guide only.
Operators' daily control numbers may be compared to the supplied
information as an operator, equipment and analyzer calibration check. Daily
control data should typically test in the range of data points provided.
The control point for the MRL Beta-lactam and Tetracycline 2 Minute Test is always zero.