



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

TYPICAL COUNTS for

Charm MRL Aflatoxin M1 Quantitative Test for Milk (MRLAFMQ)

Certificate of Quality

Product Name: Charm MRL Aflatoxin M1 Quantitative Test For Milk

Lot Number: MRLAFMQ-022J

Expiration Date: April 2022

Order Codes: LF-MRLAFMQ-20K, LF-MRLAFMQ-100K, LF-MRLAFMQ-500K

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: Sam Atoui

Title: Quality Assurance Documentation Manager

Date: 09/21/2021

Note: Typical counts are provided as a performance guide only.

Operators' daily control numbers may be compared to the supplied

information as an operator, equipment and analyzer calibration check. Daily

control data should typically fall in the range of data points provided.

The control point for the MRL Aflatoxin M1 Quantitative Test (MRLAFMQ) is 33.

TYPICAL COUNTS for Charm MRL Aflatoxin M1 Quantitative Test for Milk (MRLAFMQ) 11/11/20

Nothing works like a Charm

www.charm.com