

Certificate of Analysis

Product Name:PocketSwab PlusLot Number:8073123BExpiration Date:June 2024

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

POCK-PLUS

Quality Control Test	Specification	Result
Negative Control	0 RLU	Pass
ATP Positive Control	10,000 to 40,000 RLU	Pass

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Order Codes:

Name: David R. Legg Title: Vice President, Quality Assurance Date: 08/30/2023