



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Analysis

Product Name: Peel Plate HET – Heterotrophic Count Test

Lot Number: PP-HET-01B-017B

Expiration Date: February 29, 2024

Order Code: PP-HET

Test	Performance Specification	Result
pH of rehydrated media with water	7.2 +/- 0.2	Pass

Organism Tested	Performance Specification	Result
<i>Acinetobacter</i> sp. ATCC 49139	+/- 0.2 log*	Pass
<i>Pseudomonas fluorescens</i> ATCC 13525	+/- 0.2 log*	Pass
<i>Sphingomonas</i> sp. ATCC 31853	+/- 0.2 log*	Pass

*Log difference between test and reference method R2A Agar

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg
Title: Vice President, Quality Assurance
Date: 09/20/2022