



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Analysis

Product Name: PocketSwab Plus  
Lot Number: 8042822D  
Expiration Date: March 2023  
Order Codes: POCK-PLUS

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

Quality Control Test	Specification	Result
Negative Control	0 RLU	Pass
ATP Positive Control	10,000 to 40,000 RLU	Pass

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

A handwritten signature in dark ink, appearing to read "David R. Legg".

Name: David R. Legg  
Title: Vice President, Quality Assurance  
Date: 06/15/2022

Nothing works like a Charm