

Certificate of Analysis

Product Name: PocketSwab Plus Lot Number: 8060123C

Expiration Date: May 2024

Order Codes: POCK-PLUS

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

Quality Control Test	Specification	Result
Negative Control	0 RLU	Pass
ATP Positive Control	10,000 to 40,000 RLU	Pass

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg Title: Vice President, Quality Assurance Date: 06/20/2023