



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Analysis

Product Name: PocketSwab Plus
Lot Number: 8061520D
Expiration Date: May 2021
Order Codes: ECO-PS-PLS-100, ECO-PS-PLS-25, POCK-PLUS-100, POCK-PLUS-25

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

| Quality Control Test | Specification | Result |
|----------------------|----------------------|--------|
| Negative Control | 0 RLU | Pass |
| ATP Positive Control | 10,000 to 40,000 RLU | Pass |

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

A handwritten signature in dark ink, appearing to read "David R. Legg".

Name: David R. Legg
Title: Vice President, Quality Assurance
Date: 07/06/2020

Nothing works like a Charm