## **Certificate of Analysis**

Product Name:

PocketSwab Plus

Lot Number:

8062620D

**Expiration Date:** 

May 2021

Order Codes:

ECO-PS-PLS-100, ECO-PS-PLS-25, POCK-PLUS-100, POCK-

PLUS-25

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

Quality Control Test	Specification	Result	
Negative Control	0 RLU	Pass	
ATP Positive Control	10,000 to 40,000 RLU	Pass	

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg

Title: Vice President, Quality Assurance

Date: 07/16/2020