



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Analysis

Product Name: PocketSwab Plus  
Lot Number: 8062921D  
Expiration Date: May 2022  
Order Codes: ECO-PS-PLS-100, ECO-PS-PLS-25, POCK-PLUS-100, POCK-PLUS-25

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

| Quality Control Test | Specification        | Result |
|----------------------|----------------------|--------|
| Negative Control     | 0 RLU                | Pass   |
| ATP Positive Control | 10,000 to 40,000 RLU | Pass   |

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg  
Title: Vice President, Quality Assurance  
Date: 07/19/2021

Nothing works like a Charm