

Certificate of Analysis

Product Name:

PocketSwab Plus

Lot Number:

8063021B

Expiration Date:

May 2022

Order Codes:

ECO-PS-PLS-100, ECO-PS-PLS-25, POCK-PLUS-100, POCK-

PLUS-25

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

| Quality Control Test | Specification | Result |
|----------------------|----------------------|--------|
| Negative Control | 0 RLU | Pass |
| ATP Positive Control | 10,000 to 40,000 RLU | Pass |

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg

Title: Vice President, Quality Assurance

Date: 07/21/2021