



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Analysis

Product Name: PocketSwab Plus  
Lot Number: 8083123B  
Expiration Date: July 2024  
Order Codes: POCK-PLUS

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

| Quality Control Test | Specification        | Result |
|----------------------|----------------------|--------|
| Negative Control     | 0 RLU                | Pass   |
| ATP Positive Control | 10,000 to 40,000 RLU | Pass   |

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

A handwritten signature in blue ink, appearing to read "David R. Legg".

Name: David R. Legg  
Title: Vice President, Quality Assurance  
Date: 09/26/2023