



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Analysis

Product Name: PocketSwab Plus
Lot Number: 8100421C
Expiration Date: September 2022
Order Codes: POCK-PLUS-100, POCK-PLUS-25

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

Quality Control Test	Specification	Result
Negative Control	0 RLU	Pass
ATP Positive Control	10,000 to 40,000 RLU	Pass

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg
Title: Vice President, Quality Assurance
Date: 10/25/2021

Nothing works like a Charm