

Certificate of Analysis

Product Name:

PocketSwab Plus

Lot Number:

8110321C

Expiration Date:

October 2022

Order Codes:

POCK-PLUS-100, POCK-PLUS-25

Analysis Results for novaLUM, novaLUM II and novaLUM II X:

Quality Control Test	Specification	Result	
Negative Control	0 RLU	Pass	
ATP Positive Control	10,000 to 40,000 RLU	Pass	· ·

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg

Title: Vice President, Quality Assurance

Date: 12/03/2021