



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Quality

Product Name: ROSA Multi-Antimicrobial Positive Tablet

Lot Number: QUAD-PT-003D

Expiration Date: April 2024

Order Codes: LF-QUAD-PT

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

A handwritten signature in black ink, appearing to read "Sam Atoui", is positioned above a horizontal line.

Name: Sam Atoui

Title: Quality Assurance Documentation Manager

Date: 12/13/2022



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Quality

Product Name: Charm Quad2 Test for Macrolides in Milk

Lot Number: QUAD2-007GG

Expiration Date: November 2023

Order Codes: LF-QUAD2

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

A handwritten signature in black ink, appearing to read "Sam Atoui", is written over a horizontal line.

Name: Sam Atoui

Title: Quality Assurance Documentation Manager

Date: 04/26/2023

**TYPICAL COUNTS for**  
**Charm QUAD2 Test for Macrolides**

See Operator's Manual for use and application.

Kit Code: LF-QUAD2

**Lot Number:      QUAD2-007GG**

Charm ROSA EZ Reader

Negative Control	-3587	3129	QUAD2 Positive Control
	-3591	2909	200 ppb Spiramycin,
	-2663	2656	80 ppb Erythromycin,
	-3222	2856	300 ppb Lincomycin, and
	-3042	3237	200 ppb Tylosin
	-3436	2903	
	-3092	2914	
	-3411	3178	
	-3732	3156	
	-3849	2842	

Note: Typical counts are provided as a performance guide only.  
Operators' daily control numbers may be compared to the supplied information as an operator, equipment and analyzer calibration check. Daily control data should typically test in the range of data points provided.  
The control point for the QUAD2 Test is always zero.