



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Quality

Product Name: PKS-ATP Positive Control Tablets

Lot Number: PKS-ATP-013020E

Expiration Date: February 2024

Order Codes: ATP-PLUS

This product is made from Sigma Adenosine 5'-triphosphate disodium salt, Product Number A2383, and contains $6.0 \pm 1.5 \times 10^3$ femto moles per tablet. Finished product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

PocketSwab Plus Positive Control results for LUM channel on novaLUM should be 10,000 to 40,000 RLU. Results for Charm II 6600 or 7600 should be 1,000 to 10,000 RLU. Results for FireFly2 should be 3,000 to 20,000 RLU.

WaterGiene Positive Control results for H2OG channel on novaLUM should be 100,000,000 to 300,000,000 RLU. Results for LUM channel on novaLUM should be 100,000 to 500,000 RLU. Results for FireFly2 should be 70,000 to 100,000 RLU.

FieldSwab Positive Control results for FieldSwab channel on novaLUM should be 5,000 to 20,000 RLU.

A handwritten signature in black ink, appearing to read "Sam Atoui", is written above a horizontal line.

Name: Sam Atoui
Title: QA Documentation Manager
Date: 10/28/2022



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Quality

Product Name: Reagent Water

Lot Number: RW-090722

Expiration Date: March 2024

Order Code: RW

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

A handwritten signature in black ink, appearing to read "Sam Atoui", is written above a horizontal line.

Name: Sam Atoui
Title: Quality Assurance Documentation Manager
Date: 10/07/2022