

***B. stearothermophilus* Spore Tablet
Colony Count Certification**

Lot number: BSTTBL-029D

Expiration date: October 31, 2023

Certified colony count: 109 x 10⁶ CFU/tablet

Certification date: January 10, 2023

Colony count data supporting certification (100 x 10⁶ CFU/tablet).

Tablet #	Colony Counts			Mean
1	104	90	92	95
2	103	97	85	95
3	103	106	101	103
4	120	127	107	118
5	93	101	116	103
6	108	98	110	105
7	136	99	120	118
8	109	131	125	122
9	113	119	113	115
10	122	107	121	117
			Mean	109

x 10⁶ CFU/tablet

Colony counts of each lot are certified to be 90 x 10⁶ - 110 x 10⁶ CFU/tablet.

The average colony count of 10 tablets will be within 10% of the certified colony count (above) for that lot.



Name: David R. Legg

Title: Vice President, Quality Assurance

Date: January 10, 2023



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Quality

Product Name: B. Stearothermophilus Spore Tablets

Lot Number: BSTTBL029D

Expiration Date: October 2023

Order Code: Spore

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

The Spore tablets contain *Geobacillus stearothermophilus* var. *calidolactis*.

A handwritten signature in black ink, appearing to read "Sam Atoui", is written above a solid horizontal line.

Name: Sam Atoui
Title: Quality Assurance Documentation Manager
Date: 01/10/2023