



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Analysis

Product Name: Peel Plate Coliform High Volume Test

Lot Number: PP-CCHV-01A-028
PP-CCHV-01B-028
PP-CCHV-02A-028
PP-CCHV-02B-028
PP-CCHV-03A-028

Expiration Date: October 31, 2024

Order Code: PP-CCHV

ISO 11133:2014 Performance Testing

Organism Tested	Performance Specification	Result
<i>Escherichia coli</i> ATCC 8739 spiked dairy product	$\geq 0.5^*$	Pass
<i>Escherichia coli</i> ATCC 25922 spiked dairy product	$\geq 0.5^*$	Pass
<i>Pseudomonas aeruginosa</i> ATCC 27853 spiked dairy product	No growth	Pass

*Productivity Ratio of test to reference method, Violet Red Bile Agar (VRBA).

For a full list of inclusivity and exclusivity organisms evaluated with this EC based formulation refer to AOAC-RI Performance Tested Method #061501.

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg
Title: Vice President, Quality Assurance
Date: 05/25/2023