

**B. stearothermophilus Spore Tablet  
Colony Count Certification**

Lot number: BSTTBL-030A

Expiration date: January 31, 2025

Certified colony count: 107 x 10<sup>6</sup> CFU/tablet

Certification date: February 9, 2024

Colony count data supporting certification (100 x 10<sup>6</sup> CFU/tablet).

Tablet #	Colony Counts			Mean
1	93	106	91	97
2	99	97	87	94
3	90	123	98	104
4	92	106	107	102
5	110	100	106	105
6	108	127	124	120
7	118	125	110	118
8	127	104	112	114
9	102	119	117	113
10	117	98	94	103
			<b>Mean</b>	<b>107</b>
				<b>x 10<sup>6</sup> CFU/tablet</b>

Colony counts of each lot are certified to be 90 x 10<sup>6</sup> - 110 x 10<sup>6</sup> CFU/tablet.

The average colony count of 10 tablets will be within 10% of the certified colony count (above) for that lot.



---

Name: David R. Legg

Title: Vice President, Quality Assurance

Date: February 9, 2024



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Quality

Product Name: B. Stearothermophilus Spore Tablets

Lot Number: BSTTBL030A

Expiration Date: January 2025

Order Code: Spore

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

The Spore tablets contain *Geobacillus stearothermophilus* var. *calidolactis*.

A handwritten signature in blue ink, appearing to read "Michael T. Wade", is written over a horizontal line.

Name: Michael T. Wade  
Title: Document Control Manager  
Date: 02/12/2024